



# 深圳中电计量测试技术有限公司

Shenzhen Zhongdian Metrology and Testing Technology Co., Ltd

## 校准证书

CALIBRATION CERTIFICATE



第 1 页, 共 3 页  
Page of

证书编号

Certificate No.

ZD202206180910

委托方

Client

上海仰邦科技股份有限公司

委托方地址

Address

上海市徐汇区钦州北路1199号88幢7楼

仪器名称

Description

大气压力传感器

型号规格

Model/Type

BX-QY

制造厂

Manufacturer

上海仰邦科技股份有限公司

出厂编号

Serial No.

CGQ-2022035426

管理编号

Asset No.

/

样品接收日期

Date of Receipt

2022-06-18

校准日期

Date of Calibration

2022-06-18

批准人 :

Approved Signatory

王小云

审核 :

Inspected by

梁金学

校准 :

Tested by

梁金学



地址 : 广东省深圳市宝安区福永街道白石厦社区东区新开发区4栋301

Add : Room 301, Building 4, New Development Area, East Zone, Baishixia Community, Fuyong Sub-District, Bao'an District, Shenzhen, Guangdong, China

电话 (Tel) : 0755-29888158

邮政编码 (Post Code) : 518103

传真 (Fax) : 0755-29796107



## 校准说明

### DIRECTIONS OF CALIBRATION

证书编号: ZD202206180910

第 2 页, 共 3 页

Certificate No.

Page of

- 1.本实验室出具的数据均可溯源至国家计量基准。  
( All data issued by ZD Test are traced to National Primary Standards.)
- 2.本结果仅对当次被测样品有效, 如有疑问请在15个工作日反馈。  
( The results is ONLY valid for the tested sample, please feedback to us within 15 working days if you have any question.)
- 3.本证书编号具有唯一性, 若替换证书, 自发出后原证书即可作废。  
( Each certificate has a unique number. If replaced the certificate, the original certificate will be invalid once the new certificate number is issued.)
- 4.证书中如有最大允许误差、判定结果, 仅供参考, 其中“P”代表“Pass”, “F”代表“Fail”。  
( In the datasheet, MPE & determination is only for reference, "P" represents "Pass" and "F" represents "Fail".)
- 5.本次校准的技术依据及获认可的能力范围, 超出范围的内容未被认可。  
( Reference documents and Accredited Scopes for Calibration, Beyond the Scope has not been accredited.):

参照JJG 1084-2013《数字式气压计检定规程》《Verification Regulation of Digital Barometers》

#### 6.本次校准使用的主要测量标准 ( Main Standards of Measurement Used in the Calibration.):

名称/型号 Description/Model	编号 Serial No.	证书号 Certificate No.	有效期至 Due Date	技术特征 Technique Character
精密压力表	2016.5.61	204333889	2022-09-17	$U_{rel}=0.10\%$ , $k=2$
数字压力校验仪	20170306	223207421	2023-03-11	0.05级, $U_{rel}=0.027\%$ , $k=2$

#### 7.校准地点、环境条件 ( Place and environmental conditions of the calibration)

地点 Place	本公司实验室	温度 Temperature	22.4 °C	相对湿度 Relative Humidity	59 %
-------------	--------	-------------------	---------	---------------------------	------

- 8.建议复校时间间隔 : 12 个月, 送检单位可根据实际情况自主决定。  
Suggested calibration interval is 12 month or it can be altered depending on the actual usage of the user.



# 校准结果

## Result of Calibration

证书编号: ZD202206180910

Certificate No.

第 3 页, 共 3 页

Page of

### 1、外观以及一般性检查:

正常

In view of External and Generality check :

Normal

### 2、压力校准:

Calibration of the indicating :

项目	标称值	实测值	不确定度 $U$ ( $k=2$ )
Item	Nominal	Measured	Uncertainty
压强(kPa) Pressure	30	30.02	0.03
压强(kPa) Pressure	50	50.03	0.03
压强(kPa) Pressure	70	70.03	0.03
压强(kPa) Pressure	90	90.05	0.03
压强(kPa) Pressure	100	100.07	0.03
压强(kPa) Pressure	110	110.07	0.03

备注:

Notes:

1.依据(Reference document)

JJF 1059.1-2012 测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

(以下空白)

(The below is blank)